

## Product data sheet

## UHV gate valve, Series 108, DN 63 (ID 2½") Ordering No. 10836-UE24

## **Description**

CF-F 63 UNF Flange

Actuator pneumatic, double acting with position indicator

Feedthrough **Bellows** 

#### **Technical data**

 $< 5 \cdot 10^{-10} \text{ mbar Is}^{-1}$  $< 1 \cdot 10^{-9} \text{ mbar Is}^{-1}$  Valve body Leak rate

- Valve seat

1 · 10<sup>-10</sup> mbar to 1.6 bar (abs) Pressure range

Differential pressure on the gate ≤ 1.6 bar Differential pressure at opening ≤ 30 mbar  $600 \, ls^{-1}$ Conductance (molecular flow)

Cycles until first service 50 000 (unheated and under clean conditions)

≤ 250 °C open / ≤ 200 °C closed (bake-out max. 24h) Temperature Valve Body

≤ 200 °C (Maximum values: depending Actuator ≤ 80 °C on operating conditions and Position indicator

sealing materials)

50 °C h<sup>-1</sup> Heating and cooling rate

Material (main components) Valve Body AISI 304 (1.4301)

- Mechanism AISI 316L (1.4404), AISI 304 (1.4301)

- Bellows AISI 316L (1.4404, 1.4435)

Seal Bonnet

FKM (Viton®), vulcanized FKM (Viton®), NBR - Gate - Actuator

Mounting position any

Volume of pneumatic actuator  $0.081/0.0028 \text{ ft}^3$ 

4 - 7 bar / 58 - 102 psiCompressed air

min. - max. overpressure

1/8" ISO / NPT Compressed air connection

Created by: MAEM	Release date: 2013-01-16	1 of 2
Modified by:	Release date:	299091EA



# **Product data sheet**

# UHV gate valve, Series 108, DN 63 (ID 2½") Ordering No. 10836-UE24

Actuation time – closing 1 s

- opening 1 s

Weight 9 kg / 20 lbs

Behavior in case of compressed — Valve closed valve remains closed air pressure drop — Valve open undefined

Behavior in case of power failure — Valve closed depending on customer installation

Valve open

#### **Position indicator**

Type Micro switch Voltage  $\leq 50 \text{ V AC / DC}$ 

Current max. ≤ 1.2 A

2		OPEN		
5		CLOSED		
Valve/Ventil/Vanne				
(3 t )		OPEN	CLOSED	
3	1,2	connected	_	
Front view	6,5	-	connected	

Wiring diagram

Modified by:	Release date:	299091EA	
Created by: MAEM	Release date: 2013-01-16	2 of 2	